

Product Name	TOROIVD® 5G qPCR Premix with UNG		
Cat NO	QPT-200U	QC Date	Mar. 2nd , 2023
Lot NO	R23022720	Report Date	Mar. 2nd , 2023
QC Items	End products	Store at	2°C to 8°C for 2 years
According to	Enterprise standard ZS-B-ZLBZ-001	Expire Date	Feb. 26th , 2025
QC Standard	According to the SOP process document developed by ISO13485: 2016;		

QC Report

Items		QC Standard	Results
1	Package	The package is in good condition, no damage, no leakage, and the Lot number is correct.	<input checked="" type="checkbox"/> Pass
2	Appearance	Clear liquid, no precipitation, no suspended matter.	<input checked="" type="checkbox"/> Pass
3	pH	8.0-8.5	<input checked="" type="checkbox"/> Pass
4	Protein purity	The content of active protein should not be less than 98%.	<input checked="" type="checkbox"/> Pass
5	Nuclease residue	No endogenous endonuclease and exonuclease contamination.	<input checked="" type="checkbox"/> Pass
6	Batch stability	The difference of Cq value of different batches of reagents at ≥ 50 copies / reaction was not more than 0.5.	<input checked="" type="checkbox"/> Pass
7	Evaluation of amplification performance of internal control	The amplification curve was normal and S-shaped.	<input checked="" type="checkbox"/> Pass
		Amplification efficiency was 85% -115%, R^2 value ≥ 0.98	<input checked="" type="checkbox"/> Pass
		LOD is ≤ 50 copies	<input checked="" type="checkbox"/> Pass
8	Contamination removal capacity	Effectively remove the contamination of U-PCR products of Cq value ≥ 35 .	<input checked="" type="checkbox"/> Pass



QC Inspector		QA Manager	
---------------------	--	-------------------	--